

# REPORT No 11424

*Date of issue: November 24, 2025*

*Status: FINAL REPORT*

\*\*\*\*\*

## ASTM F1862

### RESISTANCE OF MEDICAL FACE MASKS TO PENETRATION BY SYNTHETIC BLOOD (HORIZONTAL PROJECTION OF FIXED VOLUME AT A KNOWN VELOCITY)

### Program: SQ-5053.V1

\*\*\*\*\*

This document is issued by the Company subject to its Terms and Conditions, available on request or accessible at <https://www.ptsouthquality.com/terms-and-conditions>. The Company's sole responsibility is to its Client and this document does not exonerate parties to a transaction from exercising all their rights and obligations under transaction documents. This document cannot be reproduced except in full, without prior written approval of the Company. Any unauthorized alteration, forgery or falsification of the content or appearance of this document is unlawful and offenders may be prosecuted to the fullest extent of the law.

Copyright © 2024 South Quality, Buenos Aires, ARGENTINA



<b>Prepared by:</b>	<b>Reviewed by:</b>	<b>Approved by:</b>
<b>Berenice Ferrel</b> Assistant Technician	<b>Lic. Esther Casas</b> Physics expert	<b>Eng. Emiliano Medina</b> Quality Assurance Lead

# TABLE OF CONTENTS

<b>1. FOREWORD</b>	<b>3</b>
<b>2. ORGANIZATION</b>	<b>3</b>
<b>3. OBJECTIVE</b>	<b>3</b>
<b>4. PARTICIPANT</b>	<b>3</b>
<b>5. HOMOGENEITY</b>	<b>4</b>
<b>6. SAMPLE INFORMATION</b>	<b>4</b>
<b>7. IMAGES</b>	<b>5</b>
<b>8. ASSIGNED VALUES</b>	<b>6</b>
<b>9. PARTICIPANT RESULTS</b>	<b>6</b>
<b>10. STATISTICS</b>	<b>6</b>
<b>11. EVALUATION OF PERFORMANCE</b>	<b>7</b>
<b>12. CONCLUSIONS</b>	<b>7</b>
<b>APPENDICES</b>	
<b>APPENDIX A - INSTRUCTIONS</b>	<b>8</b>
<b>APPENDIX B - PARTICIPANT RESULTS (TR - Batch LPP3051)</b>	<b>43</b>
<b>APPENDIX C - PARTICIPANT RESULTS (TR - Batch LPP3068)</b>	<b>46</b>

## 1. FOREWORD

This report summarizes the results of the SQ-5053.V1 proficiency testing program on the determination of the resistance of medical face masks to penetration of synthetic blood. This program is conducted in a bilateral format, following the A.3.3 classification of the ISO 17043 standard ("Split-sample testing schemes").

South Quality conducted the testing program in October 2025 with the aim of assessing the laboratory's ability to competently perform the designated tests.

## 2. ORGANIZATION

Program Coordinator: Lic. Esther Casas  
 Assistant Technicians: Berenice Ferrel  
 Statistic: Lic. Manuel Tozaki  
 Supervision: Eng. Emiliano Medina

## 3. OBJECTIVE

The objective of this proficiency testing program is to determination of the resistance of medical face masks to penetration of synthetic blood, using the following standard:

Standard
ASTM F1862 / F1862M - 24

To verify this, medical face mask samples have been selected.

Participants in this program have not been previously informed about the expected behavior of the samples they receive.

## 4. PARTICIPANT

Company: **Owens & Minor Halyard Health**  
 Laboratory: **La Ada de Acuña**  
 Country: Mexico  
 Client ID: C117  
 Contact person: Erika Monarrez  
 Ingeniero de Calidad I  
[erika.monarrez@owens-minor.com](mailto:erika.monarrez@owens-minor.com)

## 5. HOMOGENEITY

Several batches were prepared identically by the staff at South Quality.

Subsequently, a homogeneity study was conducted with an ISO 17025 accredited laboratory.

The control process followed ISO 33405: 2024, clauses 7.4.1.1 / 7.4.1.2. Stratified random sampling was applied, and samples were selected using random number generation software.

The results of this test are presented below:

Size of each batch: **1000 units**

Tested samples from each batch: **210 units**

DETERMINATION	HOMOGENEITY OF RESULTS IN THE ANALYZED SAMPLES (LEVEL 3 - 160 mmHG)		
	BATCH: LPP3051	BATCH: LPP3052	BATCH: LPP3053
SYNTHETIC BLOOD PENETRATION	YES	YES	NO

Size of each batch: **1000 units**

Tested samples from each batch: **210 units**

DETERMINATION	HOMOGENEITY OF RESULTS IN THE ANALYZED SAMPLES (LEVEL 3 - 160 mmHG)		
	BATCH: LPP3066	BATCH: LPP3067	BATCH: LPP3068
SYNTHETIC BLOOD PENETRATION	YES	NO	YES

Samples for this program are taken from the selected batches identified as **LPP3051** and **LPP3068**.

For the indicated batches, the values determined in the homogeneity study are utilized as the assigned values.

The analysis of the test data indicated that the selected samples exhibited sufficient homogeneity for the program. Therefore, the results of participants identified as outliers cannot be attributed to sample variability.

## 6. SAMPLE INFORMATION




The following samples were sent for testing:

Batch:	LPP3051
Sample ID:	09
Characteristics:	Medical face mask - 35 units

Batch:	LPP3068
Sample ID:	06
Characteristics:	Medical face mask - 35 units

## 7. IMAGES

### SAMPLES



## 8. ASSIGNED VALUES

BATCH	SYNTHETIC BLOOD PENETRATION - FAILURES <sup>(1)</sup> -	
	( AVG ) <sup>(2)</sup>	SD
LPP3051	29	1.4
LPP3068	5	1.2

<sup>(1)</sup> Number of failures detected in samplings of 35 units

<sup>(2)</sup> Average number of failures found in 6 samples (210 units)

## 9. PARTICIPANT RESULTS (SEE APPENDICES)

CODE	SYNTHETIC BLOOD PENETRATION ( FAILURES )
LPP3051-09	27
LPP3068-06	6

## 10. STATISTICS

The results must be treated as quantitative.

The comparison is made according B.3.1.3 of ISO 17043 and the appropriate technique is to compare participant results with the assigned values. The results can be compare using percent difference **z score**.

$$z = \frac{x - X}{\hat{\sigma}}$$

$x$  is the participant's result

$X$  is the assigned value

$\hat{\sigma}$  is the standard deviation

The performance evaluation of each sample is carried out with the following criteria:

$|z| \leq 2.0$  indicates "satisfactory" performance and generates no signal;

$2.0 < |z| < 3.0$  indicates "questionable" performance and generates a warning signal;

$|z| \geq 3.0$  indicates "unsatisfactory" performance and generates an action signal;

## 11. EVALUATION OF PERFORMANCE

BATCH	SYNTHETIC BLOOD PENETRATION - FAILURES -		z score	PERFORMANCE RESULT
	PARTICIPANT RESULT	ASSIGNED VALUE		
LPP3051	27	29	1.4	SATISFACTORY
LPP3068	6	5	0.8	SATISFACTORY

## 12. CONCLUSIONS

The overall performance on this **SQ-5053.V1** program from the participant laboratory **OWENS & MINOR HALYARD HEALTH - LA ADA DE ACUÑA**, is **SUFFICIENT** based on expected results.

The criteria used for the evaluation of the overall performance is the following:

- **SUFFICIENT** performance: No unsatisfactory/questionable results were obtained.
- **ALMOST SUFFICIENT** performance: No unsatisfactory results were obtained, but one questionable result was found.
- **INSUFFICIENT** performance: An unsatisfactory result or two questionable results were obtained.

# APPENDIX A

## INSTRUCTIONS



# INSTRUCTIONS

<b>PROGRAM:</b>	Resistance of medical face masks to penetration by synthetic blood (Horizontal projection of fixed volume at a known velocity)
<b>CODE:</b>	SQ-5053
<b>VERSION:</b>	1
<b>STANDARD:</b>	ASTM F1862
<b>COORDINATOR:</b>	Lic. Esther Casas ( <a href="mailto:ecasas@ptsouthquality.com">ecasas@ptsouthquality.com</a> )

**1 - General**

This document serves as a guide for managing the results of the **SQ-5053.V1** program.

**2 - Standard**

**ASTM F1862 / F1862M - 24**

**3 - Tests involved**

TEST
Determination of the resistance of medical face masks to penetration of synthetic blood (Level 3 – 160mmHG)

**4 - Samples**


CODE	SAMPLE	QUANTITY
LPP3051-09	Medical face mask	35
LPP3068-06	Medical face mask	35

**5 - Notes**


- a) Being a bilateral program, there is no deadline for submitting results.
- b) The participant must submit the results using the usual report employed by their laboratory.
- c) The samples are to be handled as routine lab samples, with all testing, documentation, and reporting adhering to ASTM F1862.
- d) Samples must be retained until the end of the program, which concludes with the submission of the final report.
- e) To review the results, test images would be appreciated. Images can be attached at the end of this document or sent by email.

**PHOTOGRAPHS**

1 ID 09:



2 ID 09:




3 ID 09:

DSQ-015


- REV 06 -

SQ-5053.V1

3 de 35




4 ID 09:




5 ID 09:



6 ID 09:




7 ID 09:




8 ID 09:




9 ID 09:




10 ID 09:




11 ID 09:



12 ID 09:




13 ID 09:



14 ID 09:




15 ID 09:




16 ID 09:




17 ID 09:




18 ID 09:




19 ID 09:




20 ID 09:




21 ID 09:




22 ID 09:




23 ID 09:



24 ID 09:




25 ID 09:



26 ID 09:




27 ID 09:




28 ID 09:




29 ID 09:




30 ID 09:




31 ID 09:



32 ID 09:




33 ID 09:




34 ID 09:




35 ID 09:




1 ID 06:



2 ID 06:




3 ID 06:



4 ID 06:



5 ID 06:



6 ID 06:



7 ID 06:




8 ID 06:




9 ID 06:



10 ID 06:




11 ID 06:




12 ID 06:



13 ID 06:




14 ID 06:



15 ID 06:



16 ID 06:



DSQ-012 - REV 06 -


SQ-5053.V1

26 de  
35

17 ID 06:



18 ID 06:



19 ID 06:


DSQ-012 - REV 06 -

SQ-5053.V1

27 de  
35




20 ID 06:




21 ID 06:



22 ID 06:



23 ID 06:




DSQ-012 - REV 06 -


SQ-5053.V1

29 de  
35


24 ID 06:



25 ID 06:




26 ID 06:



27 ID 06:




28 ID 06:



29 ID 06:




30 ID 06:




31 ID 06:




32 ID 06:




33 ID 06:



34 ID 06:



35 ID 06:



# APPENDIX B

## PARTICIPANT RESULTS (TR - BATCH LPP3051)

Document Number: FORM-28782 Revision: 3 Effective Date: Jun 20, 2025  
 Document Title: Laboratorio Datos Splash Test

Fecha / Date:	31-OCTUBRE-2025	Operador / Operator:	CONCEPCION ALAMILLA
Código de Producto / Product Code:	SQ-5053.V1 ID:09	Nivel de Inspección / Level Inspection:	N/A
Lote / Lot:	LPP3051	AQL:	N/A
Revision / Review:	0	Criterio de aceptacion / Acceptance criterion:	N/A
Nivel Protección / Protection Level:	3	Total de Muestras / Total Samples:	35
Maquina / Machine:	N/A	Lote y expiracion tinta / Lote and ink expiration:	2509078/30-SEPTIEMBRE-2025

### Prueba Método Preparación/Test Preparation Method

Acondicionamiento de Humedad requerido / Humidity Conditioning Required 85±5%:	Temperatura requerida en cabina / Required Chamber temperature 21 ±5°C	Tiempo de condicionamiento Mínimo Requerido / Minimum conditioning time required > 4 horas/hours
85.5	23	Promedio de Horas en Acondicionamiento / Average Hour conditioning
		21.97

### Informacion de Equipo /Equipment information

	Numero Control / Number control	Vencimiento de Calibracion / Calibration due
Cabina de acondicionamiento / Conditioning Chamber:	EQP-23027	19 AGOSTO 2026
Maquina Splash / Splash Machine:	EQP-32203	19 AGOSTO 2026
Bascula / Weighing:	EQP-12542	19 AGOSTO 2026

### Presion Requerida de Acuerdo al Nivel /Pressure required according to the level (Seleccione uno / choose one):

Nivel /Level 3	<input checked="" type="checkbox"/>	160 mmHg	Nivel /Level 2	<input type="checkbox"/>	120 mmHg	Nivel /Level 1	<input type="checkbox"/>	80 mmHg
----------------	-------------------------------------	----------	----------------	--------------------------	----------	----------------	--------------------------	---------

Total de muestras / Total of samples	35	Pasa / Pass	8	Falla / Fail	27
--------------------------------------	----	-------------	---	--------------	----

### Notas/Notes:

LOTE ACEPTADO

### Verificar todo lo aplicable / Check everything applicable

X	Todas las muestras fueron realizadas en el primer minuto despues de removerlo de las condiciones controladas de acuerdo a ASTM F1862 sec. 10.1 / All samples were made in the first minute after removing it from the controlled conditions according to ASTM F1862 sec. 10.1.
N/A	Area objetivo: en el area de sellado (PCM: donde se une el pico de pato), de acuerdo al ASTM F1862 section 9.1.1 and 9.1.2 / Target area: in the sealing area (PCM: where the duckbill meets), according to ASTM F1862 section 9.1.1 and 9.1.2.
X	Distancia desde la punta de la canula hasta el objetivo de disparo debe ser 12 pulgadas, de acuerdo con ASTM F1862 section 13.2.4 / Distance from the tip of the cannula to the shooting target should be 12 inches, in accordance with ASTM F1862 section 13.2.4.
X	Uso de algodón absorbente puede ser usado para la confirmacion de falla , de acuerdo a ASTM F1862 section 12.4.1 / Use of absorbent cotton may be used for failure confirmation, according to ASTM F1862 section 12.4.1.

Nombre y Sello de Op. Lab / Name & Stamp  
Op. Lab:

CONCEPCION ALAMILLA



Fecha / Date: 31-OCTUBRE-2025

Nombre y Sello Aprobador / Name & Stamp  
Approver:

*Erika Gonzalez Lopez*

Fecha / Date: 31-OCTUBRE-2025

THIS DOCUMENT CONTAINS CONFIDENTIAL Owens Minor Information. Printed copies are only valid on the date printed unless controlled per a manual document control procedure and stamped as such.  
Page 1 of 3

Document Number: FORM-28782		Revision: 3		Effective Date: Jun 20, 2025						
Document Title: Laboratorio Datos Splash Test										
Tiempo de condicionamiento Mínimo Requerido/Minimum conditioning time required > 4 horas/hours				Resultado de la Prueba/Testing Results						
# Muestra/Sample	Entrada/Entrance:	Salida/Exit:	Horas condicionadas / Conditioning hours:	Numero de caja / Number Box:	Pass=P Falli=F Pase=P Falla=F	Body=B Seam=S Cuerpo=B Sellado=S	2 ml Revision/Review (g)	Tiempo de disparo / Shooting Time (seg)	Realizado por / Made by:	Fecha /Date:
1	5:10:00 PM 30-Oct-2025	2:53 PM	21.71	ID-09	P	B	2.01	0.57	Concepcion Alamilla	31-Octubre-2025
2		2:54 PM	21.73		P	B				
3		2:55 PM	21.75		F	B				
4		2:56 PM	21.76		F	B				
5		2:57 PM	21.78		F	B				
6		2:58 PM	21.80		F	B				
7		2:59 PM	21.81		P	B				
8		3:00 PM	21.83		P	B				
9		3:01 PM	21.85		F	B				
10		3:02 PM	21.86		P	B				
11		3:03 PM	21.88		F	B				
12		3:04 PM	21.90		F	B				
13		3:05 PM	21.91		F	B				
14		3:06 PM	21.93		F	B				
15		3:07 PM	21.95		F	B				
16		3:08 PM	21.96		F	B	2.02	0.57		
17		3:09 PM	21.98		P	B				
18		3:10 PM	22.00		F	B	2.04	0.57		
19		3:11 PM	22.01		F	B				
20		3:12 PM	22.03		F	B				
21		3:13 PM	22.05		F	B				
22		3:14 PM	22.06		F	B				
23		3:15 PM	22.08		F	B				
24		3:16 PM	22.10		F	B				
25		3:17 PM	22.11		P	B				
26		3:18 PM	22.13		F	B				
27		3:19 PM	22.15		F	B				
28		3:20 PM	22.16		P	B				
29		3:21 PM	22.18		F	B				
30		3:22 PM	22.20		F	B				
31		3:23 PM	22.01		F	B				
32		3:24 PM	22.03		F	B				
33		3:25 PM	22.05		F	B				
34		3:26 PM	22.06		F	B				
35		3:27 PM	22.08		F	B				
36										
37										
38										
39										
40										
41										
42										
43										
44										
45										
46										
47										
48										
49										
50										

Document Number: FORM-28782

Revision: 3

Effective Date: Jun 20, 2025

Document Title: Laboratorio Datos Splash Test

Calibración antes del Orificio / Calibration difference before the Orifice			
	0.5 seg. (g)	1.5 seg (g)	Difference b/t 0.5s & 1.5s (g)
1	1.86	5.41	3.55
2	1.84	5.41	3.57
3	1.87	5.36	3.49
Average	1.86	5.39	3.54

Calibración despues del Orificio / Calibration difference after the Orifice				
Sample(s)	0.5 sec Sprut (g)	1.5 sec Sprut (g)	Difference b/t 0.5s & 1.5s (g)	% Difference
1	1.76	5.17	3.41	96%
2	1.79	5.16	3.37	94%
3	1.76	5.19	3.43	98%
Average	1.77	5.17	3.40	96.2%

Diferencia calibración antes del Orificio / Diferencia calibración despues del Orificio / Calibration difference before the Orifice / Calibration difference after the Orifice

Si el porcentaje de la diferencia esta entre 95 % y 102%, proceda a establecer el tiempo / If the percentage of the difference is between 95% and 102%, proceed to set the time.

96.2%

Si la diferencia NO esta dentro de este rango, vuelva a realizar el proceso de calibración / If the difference is NOT within this range, perform the calibration process again.

Calibración del temporizador de válvula (detrás de la placa) / Valve Timer Calibration (Behind Plate)		
#	splash gr.	
1	2.01	Ajuste el tiempo hasta que 2 g de líquido de prueba pasen a través del Orificio en tres ocasiones / Adjust the time until 2 g of test liquid passes through the Orifice on three occasions.
2	2.02	Registre el tiempo de duración para obtener los dos 2 g / Record the duration time to obtain the two 2 g
3	2	El peso debe estar entre el rango de 2.01 ± 0.04g por cada disparo individual / The weight should be within the range of 2.01 ± 0.04g for each individual shot.

Temporizador de válvula (seg.) para Sprut de 2 ml / Valve Timer (sec) for 2mL Sprut	0.57
---	------

16 Verificación de calibración de muestras (detrás de la placa) Entrega de 2 ml / 16 Speciment Calibration Check (Behind Plate) 2 mL Delivery					
Check	Sprut (g)	Comments (Pass/Fail)	Check	Sprut (g)	Comments (Pass/Fail)
1	2.01	pass	11		
2	2.02	pass	12		
3	2.04	pass	13		
4			14		
5			15		
6			16		
7			17		
8			18		
9			19		
10			20		

Lab. Condiciones / Conditioning Lab.	Inicio de pruebas / start of testing	Termino de pruebas / finished testing	Comentario / Comments
Temp (°C)	22.5	23	
Humedad (%RH)	41.6	42.6	

Nombre y Sello de Op. Lab / Name & Stamp Op. Lab:

CONCEPCION ALAMILLA



Fecha / Date: 31-OCTUBRE-2025

Nombre y Sello Aprobador / Name & Stamp Approver:

*Edu Hovaresca*

Fecha / Date: 31-OCTUBRE-2025

# APPENDIX C

## PARTICIPANT RESULTS (TR - BATCH LPP3068)

Document Number: FORM-28782      Revision: 3      Effective Date: Jun 20, 2025  
 Document Title: Laboratorio Datos Splash Test

Fecha / Date:	31-OCTUBRE-2025	Operador / Operator:	CONCEPCION ALAMILLA
Código de Producto / Product Code:	SQ-5053.V1 ID:06	Nivel de Inspección / Level Inspection:	N/A
Lote / Lot:	LPP3068	AQL:	N/A
Revision / Review:	0	Criterio de aceptación / Acceptance criterion:	N/A
Nivel Protección / Protection Level:	3	Total de Muestras / Total Samples:	35
Maquina / Machine:	N/A	Lote y expiracion tinta / Lote and ink expiration:	2509078/30-SEPTIEMBRE-2025

**Prueba Método Preparación/Test Preparation Method**

Acondicionamiento de Humedad requerido / Humidity Conditioning Required 85±5%:	Temperatura requerida en cabina / Required Chamber temperature 21 ±5°C	Tiempo de acondicionamiento Mínimo Requerido / Minimum conditioning time required > 4 horas/hours
85.7	23	Promedio de Horas en Acondicionamiento / Average Hour conditioning
		22.78

**Informacion de Equipo /Equipment information**

	Numero Control / Number control	Vencimiento de Calibracion / Calibration due
Cabina de acondicionamiento / Conditioning Chamber:	EQP-23027	19 AGOSTO 2026
Maquina Splash / Splash Machine:	EQP-32203	19 AGOSTO 2026
Bascula / Weighing:	EQP-12542	19 AGOSTO 2026

**Presion Requerida de Acuerdo al Nivel /Pressure required according to the level (Seleccione uno / choose one):**

Nivel /Level 3	<input type="checkbox"/>	160 mmHg	Nivel /Level 2	<input checked="" type="checkbox"/>	120 mmHg	Nivel /Level 1	<input checked="" type="checkbox"/>	80 mmHg
----------------	--------------------------	----------	----------------	-------------------------------------	----------	----------------	-------------------------------------	---------


Total de muestras / Total of samples	35	Pasa / Pass	29	Falla / Fail	6
--------------------------------------	----	-------------	----	--------------	---

**Notas/Notes:**  
 LOTE ACEPTADO

**Verificar todo lo aplicable / Check everything applicable**

X	Todas las muestras fueron realizadas en el primer minuto despues de removerlo de las condiciones controladas de acuerdo a ASTM F1862 sec. 10.1 / All samples were made in the first minute after removing it from the controlled conditions according to ASTM F1862 sec. 10.1.
N/A	Area objetivo: en el area de sellado (PCM: donde se une el pico de pato), de acuerdo al ASTM F1862 section 9.1.1 and 9.1.2 / Target area: in the sealing area (PCM: where the duckbill meets), according to ASTM F1862 section 9.1.1 and 9.1.2.
X	Distancia desde la punta de la canula hasta el objetivo de disparo debe ser 12 pulgadas, de acuerdo con ASTM F1862 section 13.2.4 / Distance from the tip of the cannula to the shooting target should be 12 inches, in accordance with ASTM F1862 section 13.2.4.
X	Uso de algodón absorbente puede ser usado para la confirmacion de falla , de acuerdo a ASTM F1862 section 12.4.1 / Use of absorbent cotton may be used for failure confirmation, according to ASTM F1862 section 12.4.1.

Nombre y Sello de Op. Lab / Name & Stamp Op. Lab: CONCEPCION ALAMILLA  Fecha / Date: 31-OCTUBRE-2025

Nombre y Sello Aprobador / Name & Stamp Approver:  Fecha / Date: 31-OCTUBRE-2025

THIS DOCUMENT CONTAINS CONFIDENTIAL Owens Minor Information. Printed copies are only valid on the date printed unless controlled per a manual document control procedure and stamped as such.  
 Page 1 of 3

Document Number: FORM-28782		Revision: 3		Effective Date: Jun 20, 2025						
Document Title: Laboratorio Datos Splash Test										
Tiempo de condicionamiento Mínimo Requerido/Minimum conditioning time required > 4 horas/hours				Resultado de la Prueba/Testing Results						
# Muestra/Sample	Entrada/Entrance:	Salida/Exit:	Horas condicionadas / Conditioning hours:	Numero de caja / Number Box:	Pass=P Fall=F Pasa=P Falla=F	Body=B Seams=S Cuerpo=B Sellado=S	Z ml Revision/Review (g)	Tiempo de disparo / Shooting Time (seg)	Realizado por / Made by:	Fecha /Date:
1		3:40 PM	22.5		P	B	2.01	0.57		
2		3:41 PM	22.51		P	B				
3		3:42 PM	22.53		P	B				
4		3:43 PM	22.55		P	B				
5		3:44 PM	22.56		P	B				
6		3:45 PM	22.58		P	B				
7		3:46 PM	22.60		P	B				
8		3:47 PM	22.61		F	B				
9		3:48 PM	22.63		P	B				
10		3:49 PM	22.65		F	B				
11		3:50 PM	22.66		P	B				
12		3:51 PM	22.68		P	B				
13		3:52 PM	22.70		P	B				
14		3:53 PM	22.71		F	B				
15		3:54 PM	22.73		P	B				
16		3:55 PM	22.75		P	B	2.04	0.57		
17	5:10:00 PM	3:56 PM	22.76	ID-06	P	B			Concepcion Alamilla	31-Octubre-2025
18	30-Oct-2025	3:57 PM	22.78		P	B				
19		3:58 PM	22.80		P	B				
20		3:59 PM	22.81		P	B				
21		4:00 PM	22.83		P	B				
22		4:01 PM	22.85		P	B				
23		4:02 PM	22.86		P	B				
24		4:03 PM	22.88		P	B				
25		4:04 PM	22.90		P	B				
26		4:05 PM	22.91		P	B				
27		4:06 PM	22.93		F	B				
28		4:07 PM	22.95		P	B				
29		4:08 PM	22.96		P	B				
30		4:09 PM	22.98		F	B				
31		4:10 PM	23.00	P	B					
32		4:11 PM	23.01	P	B	2.03	0.57			
33		4:12 PM	23.03	P	B					
34		4:13 PM	23.05	F	B					
35		4:14 PM	23.06	P	B					
36										
37										
38										
39										
40										
41										
42										
43										
44										
45										
46										
47										
48										
49										
50										

Document Number: FORM-28782      Revision: 3      Effective Date: Jun 20, 2025  
 Document Title: Laboratorio Datos Splash Test

Calibración antes del Orificio / Calibration difference before the Orifice			
	0.5 seg. (g)	1.5 seg (g)	Difference b/t 0.5s & 1.5s (g)
1	1.86	5.41	3.55
2	1.84	5.41	3.57
3	1.87	5.36	3.49
Average	1.86	5.39	3.54

Calibración después del Orificio / Calibration difference after the Orifice				
Sample(s)	0.5 sec Sprut (g)	1.5 sec Sprut (g)	Difference b/t 0.5s & 1.5s (g)	% Difference
1	1.76	5.17	3.41	96%
2	1.79	5.16	3.37	94%
3	1.76	5.19	3.43	98%
Average	1.77	5.17	3.40	96.2%

Diferencia calibración antes del Orificio / Diferencia calibración después del Orificio / Calibration difference before the Orifice / Calibration difference after the Orifice  
 Si el porcentaje de la diferencia esta entre 95 % y 102%, proceda a establecer el tiempo / If the percentage of the difference is between 95% and 102%, proceed to set the time.  
 Si la diferencia NO esta dentro de este rango, vuelva a realizar el proceso de calibración / If the difference is NOT within this range, perform the calibration process again.

96.2%

Calibración del temporizador de válvula (detrás de la placa) / Valve Timer Calibration (Behind Plate)		
#	splash gr.	
1	2.01	Ajuste el tiempo hasta que 2 g de líquido de prueba pasen a través del Orificio en tres ocasiones / Adjust the time until 2 g of test liquid passes through the Orifice on three occasions.
2	2.02	Registre el tiempo de duración para obtener los dos 2 g / Record the duration time to obtain the two 2 g
3	2	El peso debe estar entre el rango de 2.01 ± 0.04g por cada disparo individual / The weight should be within the range of 2.01 ± 0.04g for each individual shot.

Temporizador de válvula (seg.) para Sprut de 2 ml / Valve Timer (sec) for 2mL Sprut      0.57

16 Verificación de calibración de muestras (detrás de la placa) Entrega de 2 ml / 16 Speciment Calibration Check (Behind Plate) 2 mL Delivery					
Check	Sprut (g)	Comments (Pass/Fail)	Check	Sprut (g)	Comments (Pass/Fail)
1	2.01	pass	11		
2	2.04	pass	12		
3	2.03	pass	13		
4			14		
5			15		
6			16		
7			17		
8			18		
9			19		
10			20		

Lab. Condiciones / Conditioning Lab.	Inicio de pruebas / start of testing	Termino de pruebas / finished testing	Comentario / Comments
Temp (°C)	22.3	22.9	
Humedad (%RH)	45.8	44.3	

Nombre y Sello de Op. Lab / Name & Stamp Op. Lab:      CONCEPCION ALAMILLA           Fecha / Date: 31-OCTUBRE-2025

Nombre y Sello Aprobador / Name & Stamp Approver:           Fecha / Date: 31-OCTUBRE-2025

**----- END OF REPORT -----**